

Refractivity of Potassium Bromide for Visible Wavelengths¹

Robert J. Spindler and William S. Rodney²

Values of the index of refraction of potassium bromide were determined independently on two high-precision spectrometers by the method of minimum deviation for 11 spectral lines at temperatures near 17°, 24°, and 27° C. The instruments are described, and the methods used in obtaining results valid in the sixth decimal place of index are discussed. Temperature coefficients of index were determined, and a small linear variation was found with respect to wavelength [1]³. With the values of dn/dt so obtained, the experimental data were reduced to a common temperature, averaged, and a four-constant Ketteler-Helmholtz dispersion equation fitted to the data and adjusted by least squares. The refractive indices of potassium bromide for the particular prism used are intermediate between the values reported by Harting [2] and Forrest [3]. Reduced to 23° C for comparison, the results for sodium light are Harting, 1.55983; National Bureau of Standards, 1.55978; and Forrest, 1.55971. Another prism measured at NBS yielded 1.55983, and an obviously defective prism gave 1.55889. Small differences in dispersion were also detected.

1. Introduction

Potassium bromide is of interest to designers of optical instruments because of its transparency in the infrared region of the spectrum and because of the possibility of using it in connection with other optical materials for correction of secondary spectrum. The need for precise values of refractive indices in this connection is well known. Although some data on refractive indices of potassium bromide in the visual spectral region have been published, the lack of agreement among investigators is considerable. The authors of this paper were fortunate in having the opportunity to measure the same sample independently on two high-precision spectrometers, thereby gaining a satisfactory degree of certitude regarding the validity of the results obtained, even in the sixth decimal place.

Considerable interest has been displayed by recent investigators [4, 5] in the thermo-optical behavior of the alkali halides, primarily owing to their relatively simple crystalline structure and the information such data afford to physicists concerning interionic bond strengths. Here again, lack of precise data on temperature coefficients has handicapped physicists in theorizing upon such matters. Such considerations have led one of the authors [1] to an investigation of the dependence of index of refraction of potassium bromide upon temperature. The data thus obtained are precise enough to merit attention over the temperature range studied and are presented herein. Such data are necessary, furthermore, when reporting sixth-decimal-place values of indices of crystals whose temperature coefficients are as large as those of potassium bromide.

Investigation of the indices of different samples of the same crystalline material in the same laboratory is always of interest because of the information provided concerning control in the manufacturing

process. Such information is vitally important to optical designers, as they must know the limits between which materials vary in order to make a proper choice for their design. In this regard, differences of dispersion between samples of the substance are particularly important. Measurements were performed upon three different samples of potassium bromide grown by the Harshaw Chemical Corporation. Two of the samples were finished into 60° prisms, and the third was finished into a 30° prism by the Perkin-Elmer Corporation. The two 60° prisms were made available to the authors by Earle K. Plyler of the Radiometry Section of the National Bureau of Standards. The prism measured upon two spectrometers at three temperatures was one of these. The 30° prism was made available to the authors by Stanley S. Ballard and the Polaroid Corporation.

2. Description of Apparatus

2.1. Watts Spectrometer

One of the spectrometers (fig. 1) was made by E. R. Watts & Son, Ltd., of London, and is based on the original National Physical Laboratory design described in detail in the literature [6]. The collimator and telescope objectives have focal lengths of 60 cm and 40 cm, and clear apertures of 9 cm and 4 cm, respectively. An auxiliary telescope, identical in every respect to the rotating telescope, and mounted on a pedestal that may be clamped to the circular base, is furnished with the instrument. It may be used for a variety of purposes, including adjustment of collimation, measurement of prism angles by the substitution method, and measurement of the error in the 90° deviation of optical squares.

The collimator slit of the instrument is provided with a device for opening and closing the slit jaws symmetrically. A graduated drum is attached to the slit-width adjusting screw in order that slit widths can be recorded and repeated when necessary. Two parallel vertical lines bisected by a horizontal line are etched upon a glass plate that is mounted im-

¹ The work described in this paper was carried out in part under the sponsorship of the U. S. Air Force and U. S. Army Signal Corps.

² Some of the results reported in this paper were submitted by William S. Rodney as partial fulfillment of the requirements for the degree of Master of Science at The Catholic University of America, Washington, D. C., 1951.

³ Numbers in brackets refer to literature reference at the end of this paper.




FIGURE 1. *Watts spectrometer.*




FIGURE 3. *Tube lengths for perfect collimation of auxiliary telescope of Watts spectrometer.*




FIGURE 2. *Tube lengths for perfect collimation of telescope of Watts spectrometer.*




FIGURE 4. *Tube lengths for perfect collimation of collimator of Watts spectrometer.*

mediately behind the slit, that is, upon the source side. In adjusting collimation and making pointings, focus is made upon these lines in order that wide slits may be used to obtain sufficient intensity and to reduce chromatic parallax, as recommended by Guild [7].

Both telescopes are provided with rotatable eyepiece micrometers, permitting rapid and accurate determination of small angles and pyramidal errors of prisms. Several eyepieces of various powers are provided, and one of the Gauss type is used for leveling and autocollimation.

A removable table containing two slides that travel in superimposed ways at right angles to each other was constructed for the prism table of the instrument. The slides are actuated by slow-motion screws and greatly facilitate centering the prism with respect to collimator and telescope objectives.

Both telescopes and the collimator are provided with a rack-and-pinion device for focusing purposes. A graduated circular drum is mounted upon the pinion gear and positions of the rack may be noted by means of the graduations. By means of a monochromator manufactured by the Carl Zeiss Co. of

Jena, fitted with a flint-glass prism as dispersing agent, several spectral lines of a cadmium-mercury FH-4 lamp were separated and used as sources to illuminate the collimator slit. All three tube lengths were adjusted for truly parallel monochromatic light of each particular wavelength by the usual method of focusing upon each other in pairs, as described by Glazebrook [8]. The drum readings were recorded, and the results are shown in figures 2, 3, and 4. With the aid of these data the collimator and telescope may be adjusted for truly parallel monochromatic light. If the prism has appreciable power, owing, for example, to curved surfaces, this will be evidenced by the modification of telescope tube length necessary for sharp focus of the collimator-slit image.




FIGURE 5. Error in the mean of opposite 90° angles as a function of circle orientation.

The microscopes are mounted on a ring that may be rotated independently around the circle. This greatly facilitates the reading of the microscopes, for their positions may be so adjusted as to eliminate the troublesome cases where the microscopes are inaccessible for certain telescope orientations. The micrometer screws make 1 revolution per minute of arc, and the drums are graduated to read half-seconds directly. The elimination of parallax was effected by proper focusing, and the tube length of each microscope was adjusted to insure proper magnification so that 5 revolutions of the drum covered the average of a few 5-min intervals ruled upon the circle.

The divided circle is ruled on the periphery of a 15-in. cylindrical drum and is graduated to 5-min intervals. A complete calibration of the circle graduations has not been made, but a test showing the error in the mean of readings for opposite angles caused by positional errors of the degree graduations has been made at 10° intervals completely around the circle. For convenience, the size of the opposite angles was chosen to be 90° . The results of this test are shown in figure 5. The maximum value of 4.2 sec is uncomfortably large, but as is more or less well known [9], the effect of positional error of the circle markings upon measurements can be satisfactorily minimized in the average, provided a systematic program of observations using proper orientations of the circle is followed.

Errors of subintervals of the degree intervals were determined by measurements on four 1° intervals near 0° , 90° , 180° , and 270° , respectively. These measurements show that there is a periodic error that repeats six times in each degree interval and has an amplitude of approximately 1 sec. The periodic nature of these errors is such that their effect will be largely minimized in the average of two microscopes by making their angular displacement $179^\circ 55'$ instead of exactly 180° . Thus in figure 6 the errors of the degree subintervals as seen under microscope A are shown by the solid line. If corresponding intervals of the circle pass under microscopes A and B simultaneously, the degree subinterval errors seen under microscope B would be identical with those seen under A, and no reduction would be possible by




FIGURE 6. Error of degree subintervals as a function of microscope reading on initial graduation of each 5-minute subinterval (Watts spectrometer).

averaging A and B. If, however, the phase relationship of the error is altered by displacing microscope B one-half period (5 min), then the errors of the degree subintervals, as they will appear under B, are shown by the dotted line in figure 6. The residual error after averaging both microscopes will be reduced to the amount shown by the heavy line.

The report of an examination by the National Physical Laboratory of the circle of a similar spectrometer manufactured by the same firm is of interest because one might suspect the general characteristics of the two circles to be similar, assuming that both were ruled upon the same engine. This report mentions a systematic error that "recurs every ten minutes of arc and which results in a mean disparity of $1.05 (\pm 0.1)$ seconds in the alternate spaces between adjacent graduations." This is adequately descriptive of the error of the degree subintervals found for the instrument in the NBS laboratory. The report further states that the main systematic error can be represented "by a periodic error in the graduations of which only half a complete period is comprised within the full circle of the scale. The beginning and end of this half-period coincides closely with the zero graduation." Designating by θ an angle measured from the scale zero, the report further states "The error is zero for $\theta = 0^\circ$, 180° , or 360° and has a maximum value of $+1.95$ seconds for $\theta = 90^\circ$ or 270° . The error is positive for all values of θ , the errors of the second half of the scale (180° to 360°) being a repetition of those of the first half without change of sign." "The rms value of the random component of the error of position of a graduation is 0.60 ± 0.07 seconds."

The test performed here on the circle of the Bureau's instrument does not yield directly comparable information, but it is of interest to estimate the results that would probably be obtained by testing the circle referred to in the NPL report in the same manner in which the circle mounted in the instrument in the Bureau's laboratory was tested. The periodicity of the systematic error is the same for both circles. In making a 90° opposite-angle test on the circle




FIGURE 7. Gaertner spectrometer.

described in the NPL report, the error would be 0 for one 90° angle as read between 315° and 45° (the other 90° angle read from 165° to 225°) and $+1.95$ sec for one 90° angle as read between 0° and 90° (the other 90° angle read between 180° and 270°). The superposed random component of $\text{rms} = \pm 0.60$ sec for line position contributes an rms random component of 0.60 sec for the mean of opposite-angle readings. If it chanced that the random component should be as large as twice the rms value for lines used when the first 90° angle is read in the vicinity of 0° to 90° , then an error of $+3.2$ sec would be found, as compared with $+4$ sec, the extreme found on the instrument tested here.

Thus it would appear that the over-all quality of the circle reported upon by the NPL may be somewhat higher than the one tested at the NBS. Nevertheless, the characteristics of these circles with respect to magnitude and periodicity of errors are very similar.

2.2. Gaertner Spectrometer

The other spectrometer (fig. 7) was made by the Gaertner Scientific Corp., Chicago, Ill. It is designated as type L-124 and is of recent design. The telescope and collimator each have focal lengths of 50 cm and a clear aperture of 5 cm. Several eyepieces of various powers are provided, including a Gaussian used for leveling and autocollimation. Slow-motion screws are provided for making small angular displacements of the telescope and the divided circle. The collimator slit is provided with an auxiliary prism to facilitate comparison of spectra. A vertical spider thread is mounted in the center of the collimator slit, so that wide slits could be used and pointings made upon the spider thread in order to lessen chromatic parallax [7]. A removable table, similar to the one described in section 2.1, is mounted upon the prism table of the instrument.

The circle, which is of stainless steel, has a diameter of 300 mm. A zone about 1 in. wide near the periphery of the circle was ground and polished optically flat by the Bureau's Optical Shop. Lines separated by 5 min of arc were ruled upon the polished zone by Benjamin L. Page in the Length Sec-




FIGURE 8. Error of degree subintervals as a function of microscope reading on initial graduation of each 5-minute subinterval (Gaertner spectrometer).

tion of the National Bureau of Standards. The errors of position of the degree graduations have not been explicitly investigated, but a test of the error of the mean of opposite 90° angles, similar to that described in section 2.1, was made. The results of this test are plotted in figure 5. The maximum value of 1.1 sec indicates that the effect of positional errors of the degree markings are much less serious, approximately one-fourth of those on the Watts instrument. The same systematic program of observations mentioned in section 2.1 was used in making all determinations on this instrument.

The results of an investigation of errors in the degree subintervals are shown in figure 8. The error of the intervals, although not so completely periodic as was the case with the circle of the Watts instrument, appears to repeat itself twice in each degree interval and has a maximum amplitude of 0.9 sec of arc. By making the angular separation of two micrometer microscopes $179^\circ 45'$ instead of exactly 180° , these errors will be largely minimized in their average, as discussed in section 2.1. Thus in figure 8 the errors as seen under microscope 1 are shown by the solid line. If microscope 2 is displaced one-half period (15 min) so that the errors seen under it are of the amounts indicated by the dotted line, the amplitude of the residual error in the average of the two will be reduced to a uniform value of 0.4 sec, as shown by the heavy line.

3. Experimental Program and Procedure

Index determinations were made at temperatures near 17° , 22° , and 27° C for 11 spectral lines. The sample was kept in a dessicator while not actually being measured. The relative humidity of the room in which the measurements were made was regulated so as not to exceed 45 percent. No deterioration of the surfaces of the sample became noticeable during the course of the experiments. Care was taken to allow the sample sufficient time to reach temperature equilibrium before making deviation measurements.

Fifteen or more hours were allowed for reaching temperature equilibrium whenever the temperature of the surroundings was changed.

All refracting-angle determinations were made by autocollimation, with the Gauss eyepieces supplied with the instruments. Measurements at each of the three temperatures failed to demonstrate any temperature dependence of the refracting angle. In order to minimize torsional errors of the cone, the procedure recommended by Tilton [9] was followed, one-half the measurements being made on the angle itself and the others on its explement.⁴ The difference between the average of these two sets of data was observed to be negligible on both spectrometers at 22° C and 27° C but amounted to about 1.5 sec on the Gaertner instrument at the lower temperature. The refracting angle was of the order of magnitude of 60°. Symmetrical use of the circle was achieved by using three different orientations displaced 60° from each other for each set of angle measurements.

Measurements of double deviation were made for each spectral line after the prismatic sample had been centered with respect to the telescope and collimator objectives. Because the double deviation was always of the order of 90°, two determinations with the circle orientations 90° apart suffice to cancel approximately the effects of positional errors of the circle rulings. At least two, and sometimes four, individual determinations of double deviation were made on each instrument for each spectral line measured.

The temperature and the humidity of the laboratory in which the measurements were made were automatically controlled. The prism temperature was assumed to be the same as that recorded by a mercury thermometer suspended directly over the prism, with the mercury bulb almost in contact. The temperature was recorded at equal time intervals during the course of the observations by an observer viewing the thermometer through a telescope. The thermometer could be read directly to the nearest tenth of a degree, and the hundredths estimated. The thermometer was calibrated by the Bureau's Thermometry Section, and the correction applied over the range used was negligible. Care was taken to shield the thermometer from sources of radiation. The light source was located in a fiberboard container, and the room was more or less darkened while the observations were being made. Fluctuations in temperature were noted because of the unavoidable proximity of the observer operating the spectrometer. The average deviation of the recorded temperatures was 0.03° C. The room temperature was adjusted to the desired value, and after sufficient time had elapsed for equilibrium conditions to prevail, the deviation measurements were made.

All indices reported in this paper were corrected to refer to dry air at 760-mm pressure (Hg), and to the air temperature listed, by use of tables and procedures described by Tilton [10]. Temperature coefficients were computed for each of the 11 wave-

lengths and plotted. A straight line was passed through these data, and a least-squares adjustment of the constants was made. With the values of dn/dt computed from this equation, (2), all observed data were adjusted to 22° C. A four-constant Ketteler-Helmholtz dispersion equation was fitted to the weighted mean of the observed data and adjusted by the method of least squares.

4. Results

Values of refractive index at 22° C were computed for wavelengths ranging from 4000 to 7100 Å from the dispersion equation

$$n^2 = 2.3618102 - 0.00058072\lambda^2 + \frac{0.02305269}{\lambda^2 - 0.0245381}, \quad (1)$$

and for temperatures ranging from 16° to 28° C from the temperature coefficient equation

$$\frac{dn}{dt} = -(31.742 + 13.21\lambda) \times 10^{-6}, \quad (2)$$

where λ is in microns. These values are listed in table 1. The tabular intervals are so chosen that sixth-decimal-place indices may be obtained by linear interpolation. Table 2 lists light sources, spectral lines, wavelengths, observed and computed indices, and their differences. Table 3 lists values of dn/dt observed and computed and their differences.

Figure 9 shows a comparison of the values reported by other investigators [2, 3, 11, 12, 13]. The ordinates represent differences between the values of other observers and values reported in table 1 of this paper. Christiansen's values, like Sprockhoff's are only for the C, D, and F lines. Gyulai's values were determined on an instrument using a physical detector. He was chiefly interested in the ultraviolet indices, and the visible data represent one extreme of his observations. The most recently reported values are those of Forrest [3] and Harting [2]. The average difference at 20° C between Forrest's values and the authors' is approximately -7×10^{-5} and is constant within $\pm 5 \times 10^{-6}$; and between Harting's values and the authors' the difference is $+4 \times 10^{-5}$. The constancy in the sixth decimal place between Forrest's values and those of the authors' indicate that the internal precision of both sets of data is very high. The constancy with respect to Harting's data is of the order of one unit in the fifth decimal place, indicating a somewhat lesser degree of precision in his work. The average difference between the values reported by Forrest and by Harting is of the order of one unit in the fourth decimal place. Both these sets of data were determined in comparatively recent times on high-precision spectrometers by using modern techniques, and the differences between the two sets of data are probably caused by varying amounts of impurities in the samples themselves.

⁴ "Explement" is defined as the angle that must be added to a given angle to produce 360 degrees.

TABLE 1. *Refractive indices of potassium bromide*

[Computed from equations 1 and 2]

Wavelength <i>A</i>	Refractive index at —												
	16° C	17° C	18° C	19° C	20° C	21° C	22° C	23° C	24° C	25° C	26° C	27° C	28° C
4000	1.591415	1.591378	1.591341	1.591304	1.591267	1.591230	1.591193	1.591156	1.591119	1.591082	1.591045	1.591008	1.590971
4010	1.591101	1.591064	1.591027	1.590990	1.590953	1.590916	1.590879	1.590842	1.590805	1.590768	1.590731	1.590694	1.590657
4020	1.590789	1.590752	1.590715	1.590678	1.590641	1.590604	1.590567	1.590530	1.590493	1.590456	1.590419	1.590382	1.590345
4030	1.590480	1.590443	1.590406	1.590369	1.590332	1.590295	1.590258	1.590221	1.590184	1.590147	1.590110	1.590073	1.590036
4040	1.590174	1.590137	1.590100	1.590063	1.590026	1.589989	1.589952	1.589915	1.589878	1.589841	1.589804	1.589767	1.589730
4050	1.589872	1.589834	1.589797	1.589760	1.589723	1.589686	1.589649	1.589612	1.589575	1.589538	1.589501	1.589464	1.589426
4060	1.589571	1.589534	1.589496	1.589459	1.589422	1.589385	1.589348	1.589311	1.589274	1.589237	1.589200	1.589162	1.589125
4070	1.589273	1.589236	1.589198	1.589161	1.589124	1.589087	1.589050	1.589013	1.588976	1.588939	1.588902	1.588864	1.588827
4080	1.588978	1.588941	1.588904	1.588866	1.588829	1.588792	1.588755	1.588718	1.588681	1.588644	1.588606	1.588569	1.588532
4090	1.588685	1.588648	1.588611	1.588573	1.588536	1.588499	1.588462	1.588425	1.588388	1.588351	1.588313	1.588276	1.588239
4100	1.588395	1.588358	1.588321	1.588283	1.588246	1.588209	1.588172	1.588135	1.588098	1.588061	1.588023	1.587986	1.587949
4110	1.588107	1.588070	1.588033	1.587996	1.587958	1.587921	1.587884	1.587847	1.587810	1.587772	1.587735	1.587698	1.587661
4120	1.587822	1.587785	1.587748	1.587711	1.587673	1.587636	1.587599	1.587562	1.587525	1.587487	1.587450	1.587413	1.587376
4130	1.587539	1.587502	1.587465	1.587428	1.587390	1.587353	1.587316	1.587279	1.587242	1.587204	1.587167	1.587130	1.587093
4140	1.587259	1.587222	1.587185	1.587148	1.587110	1.587073	1.587036	1.586999	1.586962	1.586924	1.586887	1.586850	1.586813
4150	1.586981	1.586944	1.586907	1.586870	1.586832	1.586795	1.586758	1.586721	1.586684	1.586646	1.586609	1.586572	1.586535
4160	1.586706	1.586669	1.586632	1.586595	1.586557	1.586520	1.586483	1.586446	1.586409	1.586371	1.586334	1.586297	1.586260
4170	1.586434	1.586396	1.586359	1.586322	1.586285	1.586247	1.586210	1.586173	1.586136	1.586098	1.586061	1.586024	1.585987
4180	1.586163	1.586125	1.586088	1.586051	1.586014	1.585976	1.585939	1.585902	1.585864	1.585827	1.585790	1.585753	1.585715
4190	1.585895	1.585857	1.585820	1.585783	1.585746	1.585708	1.585671	1.585634	1.585596	1.585559	1.585522	1.585485	1.585447
4200	1.585629	1.585591	1.585554	1.585517	1.585480	1.585442	1.585405	1.585368	1.585330	1.585293	1.585256	1.585219	1.585181
4210	1.585365	1.585328	1.585290	1.585253	1.585216	1.585178	1.585141	1.585104	1.585066	1.585029	1.584992	1.584954	1.584917
4220	1.585103	1.585066	1.585028	1.584991	1.584954	1.584916	1.584879	1.584842	1.584804	1.584767	1.584730	1.584692	1.584655
4230	1.584844	1.584807	1.584769	1.584732	1.584695	1.584657	1.584620	1.584583	1.584545	1.584508	1.584471	1.584433	1.584396
4240	1.584587	1.584550	1.584512	1.584475	1.584438	1.584400	1.584363	1.584326	1.584288	1.584251	1.584214	1.584176	1.584139
4250	1.584332	1.584295	1.584257	1.584220	1.584183	1.584145	1.584108	1.584071	1.584033	1.583996	1.583959	1.583921	1.583884
4260	1.584078	1.584041	1.584003	1.583966	1.583929	1.583891	1.583854	1.583817	1.583779	1.583742	1.583705	1.583667	1.583630
4270	1.583828	1.583791	1.583754	1.583716	1.583679	1.583641	1.583604	1.583567	1.583529	1.583492	1.583454	1.583417	1.583380
4280	1.583579	1.583542	1.583505	1.583467	1.583430	1.583392	1.583355	1.583318	1.583280	1.583243	1.583205	1.583168	1.583131
4290	1.583329	1.583292	1.583255	1.583217	1.583180	1.583142	1.583105	1.583067	1.583030	1.582992	1.582955	1.582917	1.582880
4300	1.583089	1.583051	1.583014	1.582976	1.582939	1.582901	1.582864	1.582827	1.582789	1.582752	1.582714	1.582677	1.582639
4310	1.582846	1.582808	1.582771	1.582733	1.582696	1.582658	1.582621	1.582584	1.582546	1.582509	1.582471	1.582434	1.582396
4320	1.582605	1.582567	1.582530	1.582492	1.582455	1.582417	1.582380	1.582343	1.582305	1.582268	1.582230	1.582193	1.582155
4330	1.582367	1.582329	1.582292	1.582254	1.582217	1.582179	1.582142	1.582105	1.582067	1.582030	1.581992	1.581955	1.581917
4340	1.582130	1.582092	1.582055	1.582017	1.581980	1.581942	1.581905	1.581868	1.581830	1.581793	1.581755	1.581718	1.581680
4350	1.581895	1.581857	1.581820	1.581782	1.581745	1.581707	1.581670	1.581633	1.581595	1.581558	1.581520	1.581483	1.581445
4360	1.581662	1.581625	1.581587	1.581550	1.581512	1.581475	1.581437	1.581399	1.581362	1.581324	1.581287	1.581249	1.581212
4370	1.581431	1.581394	1.581356	1.581319	1.581281	1.581244	1.581206	1.581168	1.581131	1.581093	1.581056	1.581018	1.580981
4380	1.581201	1.581164	1.581126	1.581089	1.581051	1.581014	1.580976	1.580939	1.580901	1.580863	1.580826	1.580788	1.580751
4390	1.580974	1.580937	1.580899	1.580862	1.580824	1.580787	1.580749	1.580711	1.580674	1.580636	1.580599	1.580561	1.580524
4400	1.580748	1.580711	1.580673	1.580636	1.580598	1.580561	1.580523	1.580485	1.580448	1.580410	1.580373	1.580335	1.580298
4410	1.580524	1.580487	1.580449	1.580412	1.580374	1.580337	1.580299	1.580261	1.580224	1.580186	1.580149	1.580111	1.580074
4420	1.580302	1.580265	1.580227	1.580190	1.580152	1.580115	1.580077	1.580039	1.580002	1.579964	1.579927	1.579889	1.579852
4430	1.580083	1.580045	1.580007	1.579970	1.579932	1.579895	1.579857	1.579819	1.579782	1.579744	1.579707	1.579669	1.579631
4440	1.579865	1.579827	1.579789	1.579752	1.579714	1.579677	1.579639	1.579601	1.579564	1.579526	1.579489	1.579451	1.579413
4450	1.579648	1.579610	1.579572	1.579535	1.579497	1.579460	1.579422	1.579384	1.579347	1.579309	1.579272	1.579234	1.579196
4460	1.579433	1.579395	1.579358	1.579320	1.579283	1.579245	1.579207	1.579169	1.579132	1.579094	1.579056	1.579019	1.578981
4470	1.579219	1.579181	1.579144	1.579106	1.579068	1.579031	1.578993	1.578955	1.578918	1.578880	1.578842	1.578805	1.578767
4480	1.579008	1.578970	1.578933	1.578895	1.578857	1.578820	1.578782	1.578744	1.578707	1.578669	1.578631	1.578594	1.578556
4490	1.578798	1.578760	1.578723	1.578685	1.578647	1.578610	1.578572	1.578534	1.578497	1.578459	1.578421	1.578384	1.578346
4500	1.578589	1.578551	1.578514	1.578476	1.578438	1.578401	1.578363	1.578325	1.578288	1.578250	1.578212	1.578175	1.578137
4510	1.578382	1.578344	1.578307	1.578269	1.578231	1.578194	1.578156	1.578118	1.578081	1.578043	1.578005	1.577968	1.577930
4520	1.578177	1.578140	1.578102	1.578064	1.578026	1.577989	1.577951	1.577913	1.577876	1.577838	1.577800	1.577762	1.577725
4530	1.577973	1.577936	1.577898	1.577860	1.577822	1.577785	1.577747	1.577709	1.577672	1.577634	1.577596	1.577558	1.577521
4540	1.577771	1.577734	1.577696	1.577658	1.577620	1.577583	1.577545	1.577507	1.577470	1.577432	1.577394	1.577356	1.577319
4550	1.577571	1.577533	1.577495	1.577457	1.577420	1.577382	1.577344	1.577306	1.577268	1.577231	1.577193	1.577155	1.577117
4560	1.577372	1.577334	1.577296	1.577258	1.577221	1.577183	1.577145	1.577107	1.577069	1.577032	1.576994	1.576956	1.576918
4570	1.577175	1.577137	1.577099	1.577061	1.577024	1.576986	1.576948	1.576910	1.576872	1.576835	1.576797	1.576759	1.576721
4580	1.576979	1.576941	1.576903	1.576865	1.576828	1.576790	1.576752	1.576714	1.576676	1.576639	1.576601	1.576563	1.576525
4590	1.576784	1.576746	1.576708	1.576670	1.576633	1.576595	1.576557	1.576519	1.576481	1.576444	1.576406	1.576368	1.576330
4600	1.576591	1.576553	1.576515	1.576477	1.576440	1.576402	1.576364	1.576326	1.576288	1.576251	1.576213	1.576175	1.576137
4610	1.576399	1.576361	1.576323	1.576285	1.576248	1.576210	1.576172	1.576134	1.576096	1.576059	1.576021	1.575983	1.575945
4620	1.576209	1.576171	1.576133	1.576096	1.576058	1.576020	1.575982	1.575944	1.575906	1.575868	1.575831	1.575793	1.575755
4630	1.576020	1.575982	1.575944	1.575907	1.575869	1.575831	1.575793	1.575755	1.575717	1.575679	1.575642	1.575604	1.575566
4640	1.575833	1.575795	1.575757	1.575720	1.575682	1.575644	1.575606	1.575568	1.575530	1.575492	1.575454	1.575417	1.575379
4650	1.575647	1.575609	1.575572	1.575534	1.575496	1.575458	1.575420	1.5					

[Computed from equations 1 and 2]

Wavelength	Refractive index at —												
	16° C	17° C	18° C	19° C	20° C	21° C	22° C	23° C	24° C	25° C	26° C	27° C	28° C
<i>A</i>													
4750	1.573861	1.573823	1.573785	1.573747	1.573709	1.573671	1.573633	1.573595	1.573557	1.573519	1.573481	1.573443	1.573405
4760	1.573869	1.573851	1.573813	1.573775	1.573737	1.573699	1.573661	1.573623	1.573585	1.573547	1.573509	1.573471	1.573433
4770	1.573871	1.573841	1.573803	1.573765	1.573727	1.573689	1.573651	1.573613	1.573575	1.573537	1.573499	1.573461	1.573423
4780	1.573873	1.573832	1.573794	1.573756	1.573718	1.573680	1.573642	1.573604	1.573566	1.573528	1.573490	1.573452	1.573414
4790	1.573882	1.573844	1.573806	1.573768	1.573730	1.573692	1.573654	1.573616	1.573578	1.573540	1.573502	1.573464	1.573426
4800	1.573915	1.572977	1.572939	1.572901	1.572863	1.572825	1.572787	1.572749	1.572711	1.572673	1.572635	1.572597	1.572559
4810	1.572850	1.572811	1.572773	1.572735	1.572697	1.572659	1.572621	1.572583	1.572545	1.572507	1.572469	1.572431	1.572393
4820	1.572686	1.572648	1.572610	1.572572	1.572534	1.572496	1.572458	1.572420	1.572382	1.572344	1.572306	1.572268	1.572230
4830	1.572522	1.572484	1.572446	1.572408	1.572370	1.572332	1.572294	1.572256	1.572217	1.572179	1.572141	1.572102	1.572064
4840	1.572360	1.572322	1.572284	1.572246	1.572208	1.572169	1.572131	1.572093	1.572055	1.572017	1.571978	1.571940	1.571902
4850	1.572199	1.572161	1.572123	1.572084	1.572046	1.572008	1.571970	1.571932	1.571894	1.571856	1.571817	1.571779	1.571741
4860	1.572039	1.572001	1.571963	1.571924	1.571886	1.571848	1.571810	1.571772	1.571734	1.571696	1.571657	1.571619	1.571581
4870	1.571881	1.571843	1.571805	1.571767	1.571728	1.571690	1.571652	1.571614	1.571576	1.571537	1.571499	1.571461	1.571423
4880	1.571723	1.571685	1.571647	1.571609	1.571570	1.571532	1.571494	1.571456	1.571418	1.571379	1.571341	1.571303	1.571265
4890	1.571567	1.571529	1.571491	1.571453	1.571414	1.571376	1.571338	1.571300	1.571261	1.571223	1.571185	1.571147	1.571109
4900	1.571411	1.571373	1.571335	1.571297	1.571258	1.571220	1.571182	1.571144	1.571106	1.571067	1.571029	1.570991	1.570953
4910	1.571257	1.571219	1.571181	1.571143	1.571104	1.571066	1.571028	1.570990	1.570951	1.570913	1.570875	1.570837	1.570799
4920	1.571103	1.571065	1.571027	1.570989	1.570950	1.570912	1.570874	1.570836	1.570798	1.570759	1.570721	1.570683	1.570645
4930	1.570949	1.570913	1.570875	1.570837	1.570799	1.570760	1.570722	1.570684	1.570646	1.570607	1.570569	1.570531	1.570493
4940	1.570801	1.570762	1.570724	1.570686	1.570647	1.570609	1.570571	1.570533	1.570494	1.570456	1.570418	1.570380	1.570341
4950	1.570651	1.570612	1.570574	1.570536	1.570498	1.570459	1.570421	1.570383	1.570344	1.570306	1.570268	1.570230	1.570191
4960	1.570502	1.570463	1.570425	1.570387	1.570349	1.570310	1.570272	1.570234	1.570195	1.570157	1.570119	1.570081	1.570042
4970	1.570353	1.570315	1.570276	1.570238	1.570200	1.570161	1.570123	1.570085	1.570046	1.570008	1.569970	1.569931	1.569893
4980	1.570206	1.570168	1.570129	1.570091	1.570053	1.570014	1.569976	1.569938	1.569899	1.569861	1.569823	1.569784	1.569746
4990	1.570060	1.570022	1.569983	1.569945	1.569907	1.569868	1.569830	1.569792	1.569753	1.569715	1.569677	1.569638	1.569600
5000	1.569915	1.569877	1.569838	1.569800	1.569762	1.569723	1.569685	1.569647	1.569608	1.569570	1.569532	1.569493	1.569455
5010	1.569771	1.569733	1.569694	1.569656	1.569618	1.569579	1.569541	1.569503	1.569464	1.569426	1.569387	1.569349	1.569311
5020	1.569628	1.569590	1.569551	1.569513	1.569475	1.569436	1.569398	1.569360	1.569321	1.569283	1.569245	1.569206	1.569168
5030	1.569486	1.569448	1.569410	1.569371	1.569333	1.569294	1.569256	1.569218	1.569179	1.569141	1.569102	1.569064	1.569026
5040	1.569344	1.569306	1.569267	1.569229	1.569191	1.569152	1.569114	1.569076	1.569037	1.568999	1.568960	1.568922	1.568884
5050	1.569204	1.569166	1.569128	1.569089	1.569051	1.569012	1.568974	1.568936	1.568897	1.568859	1.568820	1.568782	1.568744
5060	1.569060	1.569022	1.568983	1.568945	1.568907	1.568868	1.568830	1.568792	1.568753	1.568715	1.568677	1.568638	1.568600
5070	1.568927	1.568888	1.568850	1.568811	1.568773	1.568734	1.568696	1.568658	1.568619	1.568581	1.568542	1.568504	1.568465
5080	1.568793	1.568755	1.568717	1.568678	1.568640	1.568601	1.568563	1.568525	1.568487	1.568448	1.568410	1.568372	1.568333
5090	1.568653	1.568614	1.568576	1.568537	1.568499	1.568460	1.568422	1.568383	1.568345	1.568307	1.568268	1.568230	1.568191
5100	1.568517	1.568478	1.568440	1.568401	1.568363	1.568324	1.568286	1.568248	1.568209	1.568171	1.568132	1.568094	1.568055
5110	1.568382	1.568343	1.568305	1.568266	1.568228	1.568189	1.568151	1.568113	1.568074	1.568035	1.567997	1.567959	1.567920
5120	1.568248	1.568210	1.568171	1.568133	1.568094	1.568056	1.568017	1.567978	1.567940	1.567901	1.567863	1.567824	1.567786
5130	1.568115	1.568077	1.568038	1.568000	1.567961	1.567923	1.567884	1.567846	1.567807	1.567768	1.567730	1.567691	1.567653
5140	1.567983	1.567945	1.567906	1.567867	1.567829	1.567791	1.567752	1.567713	1.567675	1.567636	1.567598	1.567559	1.567521
5150	1.567852	1.567814	1.567775	1.567737	1.567698	1.567660	1.567621	1.567582	1.567544	1.567505	1.567467	1.567428	1.567390
5160	1.567721	1.567683	1.567644	1.567606	1.567567	1.567529	1.567490	1.567451	1.567413	1.567374	1.567336	1.567297	1.567259
5170	1.567592	1.567554	1.567515	1.567477	1.567438	1.567400	1.567361	1.567322	1.567284	1.567245	1.567207	1.567168	1.567130
5180	1.567464	1.567425	1.567386	1.567348	1.567309	1.567271	1.567232	1.567193	1.567155	1.567116	1.567078	1.567039	1.567000
5190	1.567335	1.567297	1.567258	1.567220	1.567181	1.567143	1.567104	1.567065	1.567027	1.566988	1.566950	1.566911	1.566872
5200	1.567209	1.567170	1.567131	1.567093	1.567054	1.567016	1.566977	1.566938	1.566900	1.566861	1.566823	1.566784	1.566745
5210	1.567082	1.567043	1.567004	1.566966	1.566927	1.566889	1.566850	1.566811	1.566773	1.566734	1.566695	1.566657	1.566618
5220	1.566957	1.566918	1.566879	1.566841	1.566802	1.566764	1.566725	1.566686	1.566648	1.566609	1.566570	1.566532	1.566493
5230	1.566832	1.566793	1.566755	1.566716	1.566677	1.566639	1.566600	1.566561	1.566523	1.566484	1.566445	1.566407	1.566368
5240	1.566708	1.566669	1.566631	1.566592	1.566553	1.566515	1.566476	1.566437	1.566399	1.566360	1.566321	1.566283	1.566244
5250	1.566585	1.566546	1.566508	1.566469	1.566430	1.566392	1.566353	1.566314	1.566276	1.566237	1.566198	1.566160	1.566121
5260	1.566463	1.566424	1.566386	1.566347	1.566308	1.566270	1.566231	1.566192	1.566154	1.566115	1.566076	1.566038	1.565999
5270	1.566341	1.566303	1.566264	1.566225	1.566186	1.566148	1.566109	1.566070	1.566032	1.565993	1.565954	1.565915	1.565877
5280	1.566220	1.566182	1.566143	1.566104	1.566065	1.566027	1.565988	1.565949	1.565911	1.565872	1.565833	1.565794	1.565756
5290	1.566100	1.566062	1.566023	1.565984	1.565945	1.565907	1.565868	1.565829	1.565791	1.565752	1.565713	1.565674	1.565636
5300	1.565981	1.565943	1.565904	1.565865	1.565826	1.565788	1.565749	1.565710	1.565672	1.565633	1.565594	1.565555	1.565517
5310	1.565863	1.565824	1.565785	1.565746	1.565707	1.565669	1.565630	1.565591	1.565552	1.565514	1.565475	1.565436	1.565397
5320	1.565745	1.565706	1.565667	1.565628	1.565589	1.565551	1.565512	1.565473	1.565434	1.565396	1.565357	1.565318	1.565279
5330	1.565628	1.565589	1.565550	1.565511	1.565473	1.565434	1.565395	1.565356	1.565317	1.565279	1.565240	1.565201	1.565162
5340	1.565512	1.565473	1.565434	1.565395	1.565357	1.565318	1.565279	1.565240	1.565201	1.565163	1.565124	1.565085	1.565046
5350	1.565396	1.565357	1.565318	1.565279	1.565241	1.565202	1.565163	1.565124	1.565085	1.565047	1.565008	1.564969	1.564930
5360	1.565282	1.565243	1.565204	1.565165	1.565127	1.565088	1.565049	1.565010	1.564971	1.564933	1.564894	1.564855	1.564816
5370	1.565167	1.565128	1.565089	1.565051	1.565012	1.564973	1.564934	1.564895	1.564856	1.564817	1.564779	1.564740	1.564701
5380	1.565054	1.565015	1.564976	1.564938	1.564899	1.564860	1.564821	1.564782	1.564743	1.564704	1.564666	1.564627	1.564588
5390	1.564941	1.564902	1.564863	1.564825	1.564786	1.564747	1.564708	1.564669	1.564630	1.564591	1.564553	1.564514	1.564475
5400	1.564829	1.564790	1.564752	1.564713	1.564674	1.564635	1.564596	1.564557	1.564518	1.564479	1.564441	1.564402	1.564363
5410	1.564718	1.564679	1.564641	1.564602	1.564563	1.564524	1.564485	1.564446	1.564407	1.564368	1.564329		

TABLE 1. *Refractive indices of potassium bromide—Continued*

[Computed from equations 1 and 2]

Wavelength <i>A</i>	Refractive index at —												
	16° C	17° C	18° C	19° C	20° C	21° C	22° C	23° C	24° C	25° C	26° C	27° C	28° C
5500.....	1.563746	1.563707	1.563668	1.563629	1.563590	1.563551	1.563512	1.563473	1.563434	1.563395	1.563356	1.563317	1.563278
5520.....	1.563537	1.563498	1.563459	1.563420	1.563381	1.563342	1.563303	1.563264	1.563225	1.563186	1.563147	1.563108	1.563069
5540.....	1.563330	1.563291	1.563252	1.563213	1.563174	1.563135	1.563096	1.563057	1.563018	1.562979	1.562940	1.562901	1.562862
5560.....	1.563127	1.563087	1.563048	1.563009	1.562970	1.562931	1.562892	1.562853	1.562814	1.562775	1.562736	1.562697	1.562658
5580.....	1.562925	1.562886	1.562846	1.562807	1.562768	1.562729	1.562690	1.562651	1.562612	1.562573	1.562534	1.562494	1.562455
5600.....	1.562726	1.562687	1.562648	1.562608	1.562569	1.562530	1.562491	1.562452	1.562413	1.562374	1.562334	1.562295	1.562256
5620.....	1.562529	1.562490	1.562451	1.562411	1.562372	1.562333	1.562294	1.562255	1.562216	1.562177	1.562137	1.562098	1.562059
5640.....	1.562334	1.562295	1.562256	1.562217	1.562177	1.562138	1.562099	1.562060	1.562021	1.561981	1.561942	1.561903	1.561864
5660.....	1.562142	1.562103	1.562064	1.562025	1.561985	1.561946	1.561907	1.561868	1.561829	1.561789	1.561750	1.561711	1.561672
5680.....	1.561952	1.561913	1.561874	1.561835	1.561795	1.561756	1.561717	1.561678	1.561639	1.561599	1.561560	1.561521	1.561482
5700.....	1.561765	1.561725	1.561686	1.561647	1.561608	1.561568	1.561529	1.561490	1.561450	1.561411	1.561372	1.561333	1.561293
5720.....	1.561579	1.561539	1.561500	1.561461	1.561422	1.561383	1.561343	1.561304	1.561264	1.561225	1.561186	1.561147	1.561107
5740.....	1.561395	1.561356	1.561316	1.561277	1.561238	1.561198	1.561159	1.561120	1.561080	1.561041	1.561002	1.560962	1.560923
5760.....	1.561213	1.561174	1.561134	1.561095	1.561056	1.561016	1.560977	1.560938	1.560898	1.560859	1.560820	1.560780	1.560741
5780.....	1.561034	1.560995	1.560955	1.560916	1.560877	1.560837	1.560798	1.560759	1.560719	1.560680	1.560640	1.560601	1.560562
5800.....	1.560856	1.560817	1.560778	1.560738	1.560699	1.560659	1.560620	1.560581	1.560541	1.560502	1.560462	1.560423	1.560384
5820.....	1.560681	1.560641	1.560602	1.560562	1.560523	1.560483	1.560444	1.560405	1.560365	1.560326	1.560286	1.560247	1.560207
5840.....	1.560508	1.560468	1.560429	1.560389	1.560350	1.560310	1.560271	1.560232	1.560192	1.560153	1.560113	1.560074	1.560034
5860.....	1.560336	1.560296	1.560257	1.560217	1.560178	1.560138	1.560099	1.560060	1.560020	1.559981	1.559941	1.559902	1.559862
5880.....	1.560166	1.560127	1.560087	1.560047	1.560008	1.559969	1.559929	1.559889	1.559850	1.559810	1.559771	1.559731	1.559692
5900.....	1.559998	1.559959	1.559919	1.559880	1.559840	1.559801	1.559761	1.559721	1.559682	1.559642	1.559603	1.559563	1.559524
5920.....	1.559832	1.559793	1.559753	1.559714	1.559674	1.559635	1.559595	1.559555	1.559516	1.559476	1.559437	1.559397	1.559358
5940.....	1.559666	1.559627	1.559588	1.559549	1.559510	1.559471	1.559431	1.559391	1.559352	1.559312	1.559273	1.559233	1.559193
5960.....	1.559506	1.559466	1.559426	1.559387	1.559347	1.559308	1.559268	1.559228	1.559189	1.559149	1.559110	1.559070	1.559030
5980.....	1.559346	1.559306	1.559267	1.559227	1.559187	1.559148	1.559108	1.559068	1.559029	1.558989	1.558949	1.558910	1.558870
6000.....	1.559187	1.559147	1.559108	1.559068	1.559028	1.558989	1.558949	1.558909	1.558870	1.558830	1.558790	1.558751	1.558711
6020.....	1.559029	1.558989	1.558950	1.558910	1.558870	1.558831	1.558791	1.558751	1.558712	1.558672	1.558632	1.558593	1.558553
6040.....	1.558874	1.558835	1.558795	1.558755	1.558715	1.558676	1.558636	1.558596	1.558557	1.558517	1.558477	1.558437	1.558398
6060.....	1.558720	1.558681	1.558641	1.558601	1.558561	1.558522	1.558482	1.558442	1.558403	1.558363	1.558323	1.558283	1.558244
6080.....	1.558568	1.558528	1.558488	1.558448	1.558409	1.558369	1.558329	1.558289	1.558249	1.558210	1.558170	1.558130	1.558090
6100.....	1.558418	1.558378	1.558338	1.558298	1.558259	1.558219	1.558179	1.558139	1.558099	1.558060	1.558020	1.557980	1.557940
6120.....	1.558269	1.558229	1.558189	1.558149	1.558110	1.558070	1.558030	1.557990	1.557950	1.557911	1.557871	1.557831	1.557791
6140.....	1.558121	1.558081	1.558041	1.558002	1.557962	1.557922	1.557882	1.557842	1.557802	1.557762	1.557723	1.557683	1.557643
6160.....	1.557975	1.557935	1.557896	1.557856	1.557816	1.557776	1.557736	1.557696	1.557656	1.557616	1.557576	1.557537	1.557497
6180.....	1.557831	1.557791	1.557752	1.557712	1.557672	1.557632	1.557592	1.557552	1.557512	1.557472	1.557432	1.557392	1.557353
6200.....	1.557689	1.557649	1.557609	1.557569	1.557529	1.557489	1.557449	1.557409	1.557369	1.557329	1.557289	1.557249	1.557209
6220.....	1.557547	1.557507	1.557467	1.557427	1.557387	1.557347	1.557307	1.557267	1.557227	1.557187	1.557147	1.557107	1.557067
6240.....	1.557407	1.557367	1.557327	1.557287	1.557247	1.557207	1.557167	1.557127	1.557087	1.557047	1.557007	1.556967	1.556927
6260.....	1.557269	1.557229	1.557189	1.557149	1.557109	1.557069	1.557029	1.556989	1.556949	1.556909	1.556869	1.556829	1.556789
6280.....	1.557132	1.557092	1.557052	1.557012	1.556972	1.556932	1.556892	1.556852	1.556812	1.556772	1.556732	1.556692	1.556652
6300.....	1.556996	1.556956	1.556916	1.556876	1.556836	1.556796	1.556756	1.556716	1.556676	1.556636	1.556596	1.556556	1.556516
6320.....	1.556863	1.556822	1.556782	1.556742	1.556702	1.556662	1.556622	1.556582	1.556542	1.556502	1.556462	1.556422	1.556381
6340.....	1.556730	1.556690	1.556650	1.556610	1.556569	1.556529	1.556489	1.556449	1.556409	1.556369	1.556329	1.556289	1.556248
6360.....	1.556598	1.556558	1.556518	1.556477	1.556437	1.556397	1.556357	1.556317	1.556277	1.556237	1.556196	1.556156	1.556116
6380.....	1.556468	1.556428	1.556388	1.556347	1.556307	1.556267	1.556227	1.556187	1.556147	1.556106	1.556066	1.556026	1.555986
6400.....	1.556339	1.556299	1.556259	1.556219	1.556178	1.556138	1.556098	1.556058	1.556018	1.555977	1.555937	1.555897	1.555857
6420.....	1.556211	1.556171	1.556131	1.556091	1.556050	1.556010	1.555970	1.555930	1.555890	1.555850	1.555810	1.555770	1.555729
6440.....	1.556085	1.556045	1.556005	1.555965	1.555924	1.555884	1.555844	1.555804	1.555764	1.555724	1.555684	1.555644	1.555603
6460.....	1.555960	1.555919	1.555879	1.555839	1.555799	1.555758	1.555718	1.555678	1.555637	1.555597	1.555557	1.555517	1.555476
6480.....	1.555836	1.555795	1.555755	1.555715	1.555675	1.555634	1.555594	1.555554	1.555513	1.555473	1.555433	1.555393	1.555352
6500.....	1.555714	1.555673	1.555633	1.555593	1.555553	1.555512	1.555472	1.555432	1.555391	1.555351	1.555311	1.555270	1.555230
6520.....	1.555592	1.555552	1.555511	1.555471	1.555431	1.555390	1.555350	1.555310	1.555269	1.555229	1.555189	1.555148	1.555108
6540.....	1.555472	1.555432	1.555392	1.555351	1.555311	1.555270	1.555230	1.555190	1.555150	1.555110	1.555069	1.555028	1.554988
6560.....	1.555353	1.555313	1.555273	1.555232	1.555192	1.555151	1.555111	1.555071	1.555030	1.554990	1.554949	1.554909	1.554868
6580.....	1.555236	1.555195	1.555155	1.555114	1.555074	1.555033	1.554993	1.554953	1.554912	1.554872	1.554831	1.554791	1.554750
6600.....	1.555119	1.555078	1.555038	1.554997	1.554957	1.554916	1.554876	1.554836	1.554795	1.554755	1.554714	1.554674	1.554633
6620.....	1.555003	1.554962	1.554922	1.554881	1.554841	1.554800	1.554760	1.554719	1.554679	1.554639	1.554598	1.554557	1.554517
6640.....	1.554888	1.554848	1.554807	1.554767	1.554726	1.554686	1.554645	1.554604	1.554564	1.554523	1.554483	1.554442	1.554402
6660.....	1.554775	1.554735	1.554694	1.554654	1.554613	1.554573	1.554532	1.554491	1.554451	1.554410	1.554370	1.554329	1.554289
6680.....	1.554662	1.554622	1.554581	1.554541	1.554500	1.554459	1.554419	1.554378	1.554338	1.554297	1.554257	1.554216	1.554176
6700.....	1.554552	1.554511	1.554470	1.554430	1.554389	1.554349	1.554308	1.554267	1.554227	1.554186	1.554146	1.554105	1.554064
6720.....	1.554442	1.554401	1.554360	1.554320	1.554279	1.554239	1.554198	1.554157	1.554117	1.554076	1.554035	1.553995	1.553954
6740.....	1.554332	1.554291	1.554251	1.554210	1.554169	1.554129	1.554088	1.554047	1.554007	1.553966	1.553925	1.553885	1.553844
6760.....	1.554224	1.554183	1.554143	1.554102	1.554061	1.554021	1.553980	1.553939	1.553899	1.553858	1.553817	1.553777	1.553736
6780.....	1.554117	1.554076	1.554036	1.553995	1.553954	1.553914	1.553873	1.553832	1.553792	1.553751	1.553710	1.553669	1.553629
6800.....	1.554010	1.553970	1										

TABLE 2. Observed and computed refractive indices of potassium bromide at 22° C

Character of source	Designation of line	Wave-lengths in air	n (observed)	n (computed)	Residual (observed minus computed)
Helium tube		μ			
Mercury arc		0.7065188	1.552449	1.552447	$+2 \times 10^{-6}$
Helium tube		.6907496	1.553207	1.553212	-5
Helium tube	He red	.6678149	1.554433	1.554430	+3
Cadmium arc	Cd red	.6438470	1.555853	1.555853	0
Helium tube	d	.5875618	1.559967	1.559966	+1
Mercury arc	e	.5460740	1.563928	1.563930	-2
Cadmium arc	Cd green	.5085824	1.568476	1.568479	-3
Mercury arc		.4916036	1.570945	1.570935	+10
Hydrogen tube	F	.4861327	1.571786	1.571789	-3
Mercury arc	g	.4358342	1.581474	1.581475	-1
Do	h_2	.4046563	1.589754	1.589753	+1

TABLE 3. Temperature coefficients of refractive index of potassium bromide

Wavelength	$\frac{dn}{dt}$ (observed)	$\frac{dn}{dt}$ (computed)	Residual (observed minus computed)
μ			
0.4047	36.75×10^{-6}	37.09×10^{-6}	-0.34×10^{-6}
.4358	37.10	37.50	-.40
.4861	38.37	38.16	+.21
.4916	37.29	38.24	-.95
.5086	39.43	38.46	+.97
.5461	39.20	38.95	+.25
.5876	40.03	39.50	+.53
.6438	39.67	40.24	-.57
.6678	41.58	40.56	+1.02
.6907	40.41	40.86	-0.45
.7065	40.83	41.07	-.24

Subsequent examination of two different samples of potassium bromide by the authors verifies the possibility of even larger differences between samples. More important, differences in dispersion between samples were indicated. Figure 10 shows plots of the differences between the two other crystals studied, and the one measured upon the two spectrometers. The values designated as P_1 represent the 60° sample upon which the most extensive series of measurements was made, P_2 is a 30° prism, and P_3 is another 60° prism. Prism P_3 gave very poor imagery, and double images of the collimator slit were detected. Upon examination, this prism appeared cracked along cleavage planes in the interior of the crystal. Single, but indistinct, images of the collimator slit appeared when the prism aperture was reduced by masking portions of the surfaces with black paper so that the radiation did not pass through the portion containing the flaws. Differences in indices amounting to as much as one unit in the third decimal place in the blue were noted for this crystal, together with a definite difference in dispersion, as indicated by the slope of the curve. The differences between P_2 and P_1 are much less pronounced, amounting to about $+6.7 \times 10^{-5}$ in the average. No difference in dispersion between these two samples was found.

5. Conclusion

The values of refractive indices in table 1 are thought to have errors not exceeding $\pm 5 \times 10^{-6}$. It should be understood that these index values are




FIGURE 9. Comparison of indices of potassium bromide reported by different observers.

Harting's data and the NBS data are given for 20° C. Christiansen's data at 18° C, Sprockhoff's at 18° C, Forrest's at 26.4° C, and Gyulai's at 48° C were reduced to their 20° C equivalents by use of the temperature coefficients in table 3.




FIGURE 10. Comparison of different crystals of potassium bromide at 23° C.

for the particular sample measured. The values reported by other observers are on the whole distributed rather symmetrically around the values reported in this paper. Systematic differences between different samples appear more important than hitherto realized. The reasons for these differences are not clearly understood but probably have to do with different amounts and types of impurities in the samples. The effect of impurities

upon the ultraviolet absorption band is well known. Melvin [14], for example, mentioned in 1931 that the transmission limits of four samples of lithium fluoride varied between 1080 and 1350Å owing to slight traces of impurities. Shifting this absorption band, of course, influences the dispersion in this neighborhood, and it appears possible in the case of potassium bromide that this influence may extend into the visible region of the spectrum. Strain will also affect the refractive index, and it is possible that prism P_3 has some internal strain. This prism was one of the first to be made in this country, and it is possible that the same degree of care was not taken in purifying the salt and removing strain by annealing as is taken today in manufacturing such crystals. It is thought that the variation demonstrated between samples P_1 and P_2 is more nearly representative of the order of magnitude of the variation to be found between different samples of potassium bromide.

6. References

- [1] W. S. Rodney, Dissertation, the Catholic University of America, Washington, D. C. (1951).
- [2] H. Harting, *Z. Instrumentenk.* **63**, 125-136 (1943).
- [3] J. W. Forrest, *J. Opt. Soc. Am.* **32**, 382 (1942).
- [4] G. N. Ramachandran, *Proc. Indian Acad. Sci. [A]* **25**, 481 (1947).
- [5] T. Radhakrishnan, *Proc. Indian Acad. Sci. [A]* **31**, 224 (1950).
- [6] J. Guild and G. W. Watts, *Proc. Opt. Conv.* **II**, 937 (1926).
- [7] J. Guild, *Proc. Phys. Soc. (London)* **29**, 311 (1916).
- [8] R. Glazebrook, *Dictionary of Applied Physics* **IV**.
- [9] L. W. Tilton, *BS J. Research* **2**, 909 (1929) RP64.
- [10] L. W. Tilton, *J. Research NBS* **14**, 393 (1935) RP393.
- [11] H. Topsoe and C. Christiansen, *Ann. chim. et phys.* **1**, 5-99 (1874).
- [12] M. Sprockhoff, *Neues Jahr. Mineral. Geol., Beilage Bd.* **18**, 117-154 (1904).
- [13] Z. Gyulai, *Z. Physik* **46**, 80 (1927).
- [14] E. Melvin, *Phys. Rev.* **37**, 1230 (1931).

WASHINGTON, D. C., July 30, 1952.